



Certificate of Traceability

Level I Laboratory Certification

This certifies that CB Laboratory, using Bio-Rad Variant II Turbo 2.0 has participated in and successfully completed NGS Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: Variant II Turbo 2.0	Calibrator Lot: 64225265, 64225266	Column Lot: 70234
Reagent Lot: 64232941, 64232938, 64309761	Calibrator Assigned Values: 5.5%, 10.2%	

Date of Certification: May 1, 2020

Certification Expires: May 1, 2021

Daniel Saults

NGSP Steering Committee Chair

Rondie R Little PhD.

NGSP Network Coordinator

Shawn Ansell

SRL director/supervisor