



Certificate of Traceability

Level I Laboratory Certification

This certifies that **CB Laboratory**, using **Bio-Rad Variant II Turbo** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Variant II Turbo	Calibrator Lot: 64107221, 64107228	Column Lot: 50285
Reagent Lot: 64119441, 64106398, 64139633	Calibrator Assigned Values: 5.4%, 10.3%	

Date of Certification: May 1, 2018

Certification Expires: May 1, 2019

David Searls

NGSP Steering Committee Chair

Randee R Little PhD.

NGSP Network Coordinator

Sharon Amulfo

SRL director/ supervisor