



Certificate of Traceability

Level I Laboratory Certification

This certifies that Covance Central Laboratory Services, using Bio-Rad Variant II Turbo 2.0 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: Variant II Turbo 2.0	Calibrator Lot: 64258031, 64258032	Column Lot: 70234 AF
Reagent Lot: 64289844, 64289399, 64309761	Calibrator Assigned Values: 5.3%, 9.9%	

Date of Certification: May 1, 2020

Certification Expires: May 1, 2021

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD.

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor