



Certificate of Traceability

Level I Laboratory Certification

This certifies that **CB Laboratory**, using **Bio-Rad Variant II Turbo 2.0** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Variant II Turbo 2.0	Calibrator Lot: 64430709, 64430710	Column Lot: 90072
Reagent Lot: 64487257, 64494997, 64499839	Calibrator Assigned Values: 5.4%, 10.1%	

Date of Certification: May 1, 2023

Certification Expires: May 1, 2024

A handwritten signature in black ink, appearing to read "David Sacks".

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read "Randie R Little PhD".

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read "Sharon Amalf".

SRL director/ supervisor